

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
-40 °C	-40.642 °C	-40.1 °C	0.542 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.212 °C	-14.9 °C	0.312 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.572 °C	-0.2 °C	0.372 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.606 °C	10.0 °C	0.394 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.5 °C	0.492 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.270 °C	40.8 °C	0.530 °C	± 0.5 °C	± 4.0E-01 °C