

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
#-40 °C	-40.642 °C	-40.1 °C	0.542 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.212 °C	-14.9 °C	0.312 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.572 °C	-0.2 °C	0.372 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.606 °C	10.0 °C	0.394 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.5 °C	0.492 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.270 °C	40.8 °C	0.530 °C	± 0.5 °C	± 4.0E-01 °C