

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
50.01 mmHg	50 mmHg	-0.01 mmHg	± 3 mmHg	±	6.6E-01 mmHg
100.02 mmHg	100 mmHg	-0.02 mmHg	± 3 mmHg	±	6.5E-01 mmHg
150.02 mmHg	150 mmHg	-0.02 mmHg	± 3 mmHg	±	6.6E-01 mmHg
200.02 mmHg	200 mmHg	-0.02 mmHg	± 3 mmHg	±	6.8E-01 mmHg
250.01 mmHg	250 mmHg	-0.01 mmHg	± 3 mmHg	±	7.5E-01 mmHg
300.02 mmHg	300 mmHg	-0.02 mmHg	± 3 mmHg	±	7.5E-01 mmHg