

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.9 °C	0.838 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	25.5 %HR	-4.763 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.488 %HR	45.3 %HR	-5.188 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	64.6 %HR	-5.848 %HR	± 5.0 %HR	± 1.9E+00 %HR