

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.071 °C	0.0 °C	0.071 °C	± 0.2 °C	± 4.0E-01 °C
2 °C	2.085 °C	2.1 °C	0.015 °C	± 0.2 °C	± 4.0E-01 °C
4 °C	4.091 °C	4.2 °C	0.109 °C	± 0.2 °C	± 4.0E-01 °C
8 °C	7.801 °C	7.9 °C	0.099 °C	± 0.2 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.2 °C	± 4.0E-01 °C