

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
0 °C	-0.071 °C	0.0 °C	0.071 °C	± 0.2 °C	± 4.0E-01 °C
2 °C	2.085 °C	2.1 °C	0.015 °C	± 0.2 °C	± 4.0E-01 °C
4 °C	4.091 °C	4.1 °C	0.009 °C	± 0.2 °C	± 4.0E-01 °C
8 °C	7.801 °C	7.9 °C	0.099 °C	± 0.2 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.2 °C	± 4.0E-01 °C