

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C / 86 °F	29.471 / 85.04 °C / °F	29.8 / 85.6 °C / °F	0.329 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.568 %HR	33.0 %HR	-1.568 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	54.118 %HR	51.4 %HR	-2.718 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.120 %HR	72.2 %HR	2.080 %HR	± 3.0 %HR	± 1.9E+00 %HR