

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.2 μm	23.7 μm	0.5 μm	± 0.7 μm	±	5.8E-01 μm
51.1 μm	52.0 μm	0.9 μm	± 1.5 μm	±	5.8E-01 μm
125.1 μm	127 μm	1.9 μm	± 3.8 μm	±	5.8E-01 μm
478.0 μm	485 μm	7.0 μm	± 14.5 μm	±	5.8E-01 μm
963 μm	972 μm	9.0 μm	± 29.1 μm	±	6.0E-01 μm