

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.099900	0.0900	± 0.6	± 4.5E-01 mm					
0 a 0.2	0.199800	0.1900	± 0.6	± 4.5E-01 mm					
0 a 0.3	0.299600	0.3800	± 0.7	± 4.5E-01 mm					
0 a 0.4	0.399900	0.0700	± 0.7	± 4.5E-01 mm					
0 a 0.5	0.499200	0.7600	± 0.8	± 4.5E-01 mm					
0 a 0.6	0.599400	0.5500	± 0.8	± 4.5E-01 mm					
0 a 0.7	0.699600	0.3300	± 0.8	± 4.5E-01 mm					
0 a 0.8	0.799300	0.6000	± 0.9	± 4.5E-01 mm					
0 a 0.9	0.899300	0.5600	± 0.9	± 4.5E-01 mm					
0 a 1	0.999000	0.8100	± 1	± 4.5E-01 mm					