

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.471 °C	30.0 °C	0.529 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	32.513 %HR	31 %HR	-1.513 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.006 %HR	47 %HR	-2.006 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.120 %HR	68 %HR	-2.120 %HR	± 5.0 %HR	± 1.9E+00 %HR