

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.001 °C	15.4 °C	0.399 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	29.832 °C	30.1 °C	0.268 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.917 °C	32.3 °C	0.383 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	19.476 %HR	22.5 %HR	3.024 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	53.3 %HR	2.546 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.051 %HR	72.2 %HR	1.149 %HR	± 5.0 %HR	± 1.9E+00 %HR