

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.801 °C        | -25.8 °C  | 0.001 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.311 °C         | -0.3 °C   | 0.011 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 8 °C              | 7.911 °C          | 8.0 °C    | 0.089 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |