

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.801 °C	-25.6 °C	0.201 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.311 °C	-0.2 °C	0.111 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.911 °C	8.1 °C	0.189 °C	± 0.5 °C	± 4.0E-01 °C