

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.591 °C	-25.0 °C	0.591 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.321 °C	0.3 °C	0.621 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.403 °C	9.0 °C	0.597 °C	± 0.5 °C	± 4.0E-01 °C