

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.591 °C        | -24.9 °C  | 0.691 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.321 °C         | 0.6 °C    | 0.921 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 8 °C              | 8.403 °C          | 8.4 °C    | -0.003 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |