

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	4.352 °C	4.6 °C	0.248 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.171 °C	15.3 °C	0.129 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.333 °C	30.6 °C	0.267 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.198 %HR	29.4 %HR	1.202 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.233 %HR	51.8 %HR	1.567 %HR	± 3 %HR	± 1.9E+00 %HR
65 %HR	64.670 %HR	66.8 %HR	2.130 %HR	± 3 %HR	± 1.9E+00 %HR