

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.032 °C	30.5 °C	0.468 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	33.044 %HR	31 %HR	-2.044 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.681 %HR	47 %HR	-2.681 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.947 %HR	73 %HR	-1.947 %HR	± 5.0 %HR	± 1.9E+00 %HR