

Certificado No. (Certificate #):

61975

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-10-06

Marca (Brand):

Starrett

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm		±	5.8E+00 µm
3.00 mm	3.00 mm	0 µm		±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm		±	5.8E+00 µm
7.00 mm	7.00 mm	0 µm		±	5.8E+00 µm
10.00 mm	10.00 mm	0 µm		±	5.8E+00 µm
14.00 mm	14.00 mm	0 µm		±	5.8E+00 µm
16.00 mm	16.00 mm	0 µm		±	5.8E+00 µm
18.00 mm	18.00 mm	0 µm		±	5.8E+00 µm
20.00 mm	20.00 mm	0 µm		±	5.8E+00 µm