

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.905 °C	2.2 °C	0.295 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.951 °C	15.2 °C	0.249 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.882 °C	30.0 °C	0.118 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.619 %HR	59.0 %HR	-0.619 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.846 %HR	67.2 %HR	-3.646 %HR	± 5.0 %HR	± 1.9E+00 %HR