

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.905 °C	2.2 °C	0.295 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.951 °C	15.2 °C	0.249 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.882 °C	29.8 °C	-0.082 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.619 %HR	58.0 %HR	-1.619 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.846 %HR	67.1 %HR	-3.746 %HR	± 5.0 %HR	± 1.9E+00 %HR