

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.905 °C	2.3 °C	0.395 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.951 °C	15.3 °C	0.349 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.882 °C	30.1 °C	0.218 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.619 %HR	58.6 %HR	-1.019 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.846 %HR	67.5 %HR	-3.346 %HR	± 5.0 %HR	± 1.9E+00 %HR