

Certificado No. (Certificate #):		61999		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-10-03		Marca (Brand):		Stereon	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		5.0 mm	4.9 mm	-100 µm	± 50 µm	± 6.2E+01 µm	
Exterior / (Exterior)		10.0 mm	9.9 mm	-100 µm	± 50 µm	± 6.2E+01 µm	
Exterior / (Exterior)		50.0 mm	49.9 mm	-100 µm	± 50 µm	± 6.2E+01 µm	
Exterior / (Exterior)		100.0 mm	99.9 mm	-100 µm	± 50 µm	± 6.2E+01 µm	
Exterior / (Exterior)		150.0 mm	149.9 mm	-100 µm	± 50 µm	± 6.2E+01 µm	
Interior / (Interior)		25.0 mm	25.1 mm	100 µm	± 50 µm	± 5.8E+01 µm	
Interior / (Interior)		100.0 mm	100.1 mm	100 µm	± 50 µm	± 5.8E+01 µm	
Profundidad / (Depth)		20.0 mm	20.2 mm	200 µm	± 50 µm	± 5.9E+01 µm	
Profundidad / (Depth)		100.0 mm	100.2 mm	200 µm	± 50 µm	± 5.9E+01 µm	