

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.062 °C         | 30.3 °C   | 0.238 °C      | ± 1.0 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.570 %HR        | 32.6 %HR  | 2.030 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.345 %HR        | 52.2 %HR  | 1.855 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.243 %HR        | 73.1 %HR  | 2.857 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |