

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.1 °C	0.038 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33.2 %HR	2.630 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.345 %HR	50.3 %HR	-0.045 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.2 %HR	-2.043 %HR	± 3.0 %HR	± 1.9E+00 %HR