

| Certificado No. (Certificate #):         |  | 62069             |           | Equipo (Instrument): |                   | Calibrador Pie de Rey       |  |
|------------------------------------------|--|-------------------|-----------|----------------------|-------------------|-----------------------------|--|
| Fecha de Calibración (Calibration Date): |  | 2025-10-10        |           | Marca (Brand):       |                   | Total                       |  |
| Tipo de Medida (Measurement Type)        |  | Patrón (Standard) | UBP (UUT) | Error (Error)        | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |  |
| Exterior / (Exterior)                    |  | 5.00 mm           | 5.01 mm   | 10 µm                | ± 30 µm           | ± 7.4E+00 µm                |  |
| Exterior / (Exterior)                    |  | 10.00 mm          | 10.01 mm  | 10 µm                | ± 30 µm           | ± 7.4E+00 µm                |  |
| Exterior / (Exterior)                    |  | 50.00 mm          | 50.02 mm  | 20 µm                | ± 30 µm           | ± 7.4E+00 µm                |  |
| Exterior / (Exterior)                    |  | 100.00 mm         | 99.97 mm  | -30 µm               | ± 30 µm           | ± 7.4E+00 µm                |  |
| Exterior / (Exterior)                    |  | 150.00 mm         | 149.97 mm | -30 µm               | ± 30 µm           | ± 7.4E+00 µm                |  |
| Interior / (Interior)                    |  | 25.00 mm          | 24.98 mm  | -20 µm               | ± 30 µm           | ± 8.2E+00 µm                |  |
| Interior / (Interior)                    |  | 100.00 mm         | 99.93 mm  | -70 µm               | ± 30 µm           | ± 8.2E+00 µm                |  |
| Profundidad / (Depth)                    |  | 20.00 mm          | 20.00 mm  | 0 µm                 | ± 30 µm           | ± 7.1E+00 µm                |  |
| Profundidad / (Depth)                    |  | 100.00 mm         | 100.00 mm | 0 µm                 | ± 30 µm           | ± 7.1E+00 µm                |  |