

Certificado No. (Certificate #):		62069		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-10-10		Marca (Brand):		Hardell	
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)		
Exterior / (Exterior)	20.00 mm	20.00 mm	0 µm	± 30 µm	± 7.7E+00 µm		
Exterior / (Exterior)	50.00 mm	50.01 mm	10 µm	± 30 µm	± 7.7E+00 µm		
Exterior / (Exterior)	100.00 mm	100.01 mm	10 µm	± 30 µm	± 7.7E+00 µm		
Exterior / (Exterior)	200.00 mm	199.99 mm	-10 µm	± 30 µm	± 7.7E+00 µm		
Exterior / (Exterior)	300.00 mm	299.99 mm	-10 µm	± 40 µm	± 7.7E+00 µm		
Interior / (Interior)	100.00 mm	99.99 mm	-10 µm	± 30 µm	± 8.2E+00 µm		
Interior / (Interior)	150.00 mm	150.03 mm	30 µm	± 30 µm	± 8.2E+00 µm		
Profundidad / (Depth)	100.00 mm	100.02 mm	20 µm	± 30 µm	± 7.1E+00 µm		
Profundidad / (Depth)	200.00 mm	200.03 mm	30 µm	± 30 µm	± 7.1E+00 µm		