

Certificado No. (Certificate #):

62072

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-10-09

Marca (Brand):

Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.050 in	0.050 in	0.000 in		±	5.8E-04 in
0.100 in	0.100 in	0.000 in		±	5.8E-04 in
0.150 in	0.150 in	0.000 in		±	5.9E-04 in
0.200 in	0.200 in	0.000 in		±	5.9E-04 in
0.400 in	0.400 in	0.000 in		±	6.4E-04 in
0.500 in	0.500 in	0.000 in		±	6.8E-04 in
0.600 in	0.600 in	0.000 in		±	7.1E-04 in
0.800 in	0.800 in	0.000 in		±	8.0E-04 in
1.000 in	1.000 in	0.000 in		±	9.1E-04 in