

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.0 °C	-0.062 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	24.0 %HR	-6.366 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.427 %HR	44.0 %HR	-6.427 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	71.0 %HR	0.859 %HR	± 5.0 %HR	± 1.9E+00 %HR