

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	15.6 °C	0.378 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	19.8 °C	-0.235 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.1 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	18.1 %HR	-2.245 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.407 %HR	52.7 %HR	2.293 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	82.6 %HR	2.439 %HR	± 3.0 %HR	± 1.9E+00 %HR