

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
15.02458 ml	15 ml	-0.02458 ml	± 0.03 ml	±	4.7E-01 µl
75.03919 ml	75 ml	-0.03919 ml	± 0.05 ml	±	2.4E+00 µl
149.88702 ml	150 ml	0.11298 ml	± 0.10 ml	±	4.6E+00 µl