

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.711 °C	29.9 °C	0.189 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	29.3 %HR	-2.712 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.644 %HR	48.8 %HR	-2.844 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.051 %HR	71.6 %HR	0.549 %HR	± 3.0 %HR	± 1.9E+00 %HR