

Certificado No. (Certificate #):

62116

Equipo (Instrument):

Reloj Palpador

Fecha de Calibración (Calibration Date):

2025-10-10

Marca (Brand):

Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.010 mm	0.01 mm	0 µm		±	5.8E+00 µm
0.030 mm	0.03 mm	0 µm		±	5.8E+00 µm
0.050 mm	0.05 mm	0 µm		±	5.8E+00 µm
0.080 mm	0.08 mm	0 µm		±	5.8E+00 µm
0.100 mm	0.10 mm	0 µm		±	5.8E+00 µm
0.500 mm	0.50 mm	0 µm		±	5.8E+00 µm
0.600 mm	0.60 mm	0 µm		±	5.8E+00 µm
0.700 mm	0.70 mm	0 µm		±	5.8E+00 µm
0.800 mm	0.80 mm	0 µm		±	5.8E+00 µm