

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.005 mm	1.010 mm	5 µm	± 50.25 µm	±	5.8E-01 µm
1.5 mm	1.506 mm	6 µm	± 75. µm	±	5.8E-01 µm
2 mm	2.006 mm	6 µm	± 100 µm	±	5.8E-01 µm
4 mm	4.008 mm	8 µm	± 200 µm	±	5.8E-01 µm
5 mm	5.001 mm	1 µm	± 250 µm	±	5.8E-01 µm