

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 29.741 °C         | 30.4 °C   | 0.659 °C      | ± 1.0 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 32.012 %HR        | 28 %HR    | -4.012 %HR    | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 51.644 %HR        | 50 %HR    | -1.644 %HR    | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 71.051 %HR        | 76 %HR    | 4.949 %HR     | ± 5.0 %HR            | ± 1.9E+00 %HR               |