

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.741 °C	30.0 °C	0.259 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	28 %HR	-4.012 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.644 %HR	48 %HR	-3.644 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.051 %HR	69 %HR	-2.051 %HR	± 5.0 %HR	± 1.9E+00 %HR