

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.741 °C	30.0 °C	0.259 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	29 %HR	-3.012 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.644 %HR	49 %HR	-2.644 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.051 %HR	71 %HR	-0.051 %HR	± 5.0 %HR	± 1.9E+00 %HR