

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.5044 ml	0.5 ml	-0.0044 ml	± 0.050 ml	±	9.6E-02 µl
1.0013 ml	1 ml	-0.0013 ml	± 0.050 ml	±	6.0E-02 µl
1.9930 ml	2 ml	0.0070 ml	± 0.080 ml	±	9.7E-02 µl
3.010577 ml	3 ml	-0.0106 ml	± 0.080 ml	±	1.3E-01 µl
3.999736 ml	4 ml	0.0003 ml	± 0.080 ml	±	1.5E-01 µl
5.010625 ml	5 ml	-0.0106 ml	± 0.125 ml	±	1.7E-01 µl