

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
1.00 ml	1 ml	-0.0010 ml	± 0.05 ml	±	7.7E-02 µl
5.00 ml	5 ml	0.0006 ml	± 0.13 ml	±	2.9E-01 µl
10.00 ml	10 ml	0.0005 ml	± 0.12 ml	±	5.7E-01 µl