

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00096 ml	1 ml	-0.00096 ml	± 0.05 ml	±	7.7E-02 µl
4.99937 ml	5 ml	0.00063 ml	± 0.13 ml	±	2.9E-01 µl
9.99955 ml	10 ml	0.00045 ml	± 0.12 ml	±	5.7E-01 µl