

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
1,000.02 µl	1000 µl	-0.0150 µl	± 8 µl	±	7.7E-02 µl
1999.93 µl	2000 µl	0.0702 µl	± 16 µl	±	1.4E-01 µl
3,000.05 µl	3000 µl	-0.0451 µl	± 24 µl	±	1.9E-01 µl
4000.06 µl	4000 µl	-0.0601 µl	± 32 µl	±	2.4E-01 µl
4999.97 µl	5000 µl	0.0251 µl	± 40 µl	±	2.9E-01 µl