

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.1001 ml	0.1 ml	-0.0001 ml	± 0.01 ml	±	5.1E-02 µl
1.0002 ml	1 ml	-0.0002 ml	± 0.01 ml	±	6.0E-02 µl
10.0061 ml	10 ml	-0.0061 ml	± 0.05 ml	±	3.2E-01 µl
25.0175 ml	25 ml	-0.0175 ml	± 0.10 ml	±	8.1E-01 µl