

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.711 °C	29.9 °C	0.189 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	29.6 %HR	-2.412 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.644 %HR	49.6 %HR	-2.044 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.051 %HR	69.4 %HR	-1.651 %HR	± 3.0 %HR	± 1.9E+00 %HR