

Tipo de Medida <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>
Profundidad / (Depth)	0 in	0.06 in	0 in	± 0.0019685 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.10 in	0 in	± 0.0019685 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.13 in	0 in	± 0.0019685 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.20 in	0 in	± 0.0019685 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.25 in	0 in	± 0.0019685 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.30 in	0 in	± 0.0019685 in	± 3.7E-03 in