

Tipo de Medida <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>
Profundidad / (Depth)	0 in	0.063 in	0.0005 in	± 0.0019 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.100 in	0.0000 in	± 0.0019 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.125 in	0.0000 in	± 0.0019 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.200 in	0.0000 in	± 0.0019 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.250 in	0.0000 in	± 0.0019 in	± 2.1E-03 in
Profundidad / (Depth)	0 in	0.300 in	0.0000 in	± 0.0019 in	± 3.7E-03 in