

Tipo de Medida <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>
Profundidad / (Depth)	1.0000 in	1.0000 in	0.0000 in	± 0.001968 in	± 2.1E-03 in
Profundidad / (Depth)	2.0000 in	2.0000 in	0.0000 in	± 0.001968 in	± 2.1E-03 in
Profundidad / (Depth)	3.0000 in	3.0000 in	0.0000 in	± 0.001968 in	± 2.1E-03 in
Profundidad / (Depth)	4.0000 in	4.0000 in	0.0000 in	± 0.002755 in	± 2.1E-03 in