

Certificado No. (Certificate #):

62166

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-10-09

Marca (Brand):

Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.050 in	0.050 in	0.000 in		±	6.7E-04 in
0.100 in	0.100 in	0.000 in		±	6.7E-04 in
0.110 in	0.110 in	0.000 in		±	6.7E-04 in
0.150 in	0.150 in	0.000 in		±	6.7E-04 in
0.200 in	0.200 in	0.000 in		±	6.8E-04 in
0.300 in	0.300 in	0.000 in		±	9.1E-04 in
0.400 in	0.399 in	-0.001 in		±	1.4E-03 in
0.500 in	0.499 in	-0.001 in		±	1.9E-03 in
1.000 in	0.997 in	-0.003 in		±	9.6E-04 in