

Certificado No. (Certificate #):62171Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-10Marca (Brand):No Indicada

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.0000 in	1.0000 in	0.0000 in	± 0.0001 in	±	5.9E-05 in
1.1000 in	1.1000 in	0.0000 in	± 0.0002 in	±	5.9E-05 in
1.2000 in	1.2000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
1.3000 in	1.3000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
1.4000 in	1.4000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
1.5000 in	1.5000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
1.6000 in	1.6000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
1.7000 in	1.7000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
1.8000 in	1.8000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
1.9000 in	1.9000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
2.0000 in	2.0000 in	0.0000 in	± 0.0002 in	±	6.0E-05 in