

Certificado No. (Certificate #):

62172

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-10-10

Marca (Brand):

-

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.050 in	0.050 in	0.000 in		±	2.9E-04 in
0.100 in	0.100 in	0.000 in		±	3.0E-04 in
0.105 in	0.105 in	0.000 in		±	4.1E-04 in
0.110 in	0.110 in	0.000 in		±	3.0E-04 in
0.150 in	0.150 in	0.000 in		±	3.1E-04 in
0.200 in	0.200 in	0.000 in		±	3.2E-04 in
0.205 in	0.205 in	0.000 in		±	4.3E-04 in
0.210 in	0.210 in	0.000 in		±	4.3E-04 in
0.250 in	0.250 in	0.000 in		±	3.4E-04 in