

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.995 °C	2.1 °C	0.105 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.900 °C	4.0 °C	0.100 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.192 °C	8.3 °C	0.108 °C	± 0.5 °C	± 4.0E-01 °C