

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.2 °C	0.138 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.478 %HR	38.0 %HR	2.522 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.427 %HR	52.6 %HR	2.173 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	72.5 %HR	2.359 %HR	± 3.0 %HR	± 1.9E+00 %HR